

Report: SSL.19B023.MB2-HW

Issued: 5 March 2019

Page: 1 of 8

Test Report:

EN 13727:2012+A2:2015

Chemical disinfectants and antiseptics – Quantitative suspension test for the evaluation of bactericidal activity in the medical area – Test method and requirements (phase 2, step 1)

Identification of the test laboratory:

Abbott Analytical Ltd
Unit 2, Hickmans Road, Birkenhead, CH41 1JH, United Kingdom

Identification of the client:

Safe Solutions (Safe4) Ltd
Wharton Green, Bostock Road, Winsford, CW7 3BD, United Kingdom

Identification of the sample:

19B/023

Name of the product:

Handwash

Batch number/reference and
expiry date (if available):

3442

Date of delivery:

18 February 2019

Storage conditions:

Room temperature in darkness

Product diluent recommended by
the manufacturer for use:

Not disclosed

Active substance(s) and their
concentrations (s) (optional):

Not disclosed

Appearance of the product:

Thick clear colourless liquid

Notes:

- 1) The test results in this report relate only to the sample(s) tested.
- 2) This test report may not be reproduced except in full, adapted, altered or used to create a derivative work, without written approval from Abbott Analytical Ltd.

Report: SSL.19B023.MB2-HW

Issued: 5 March 2019

Page: 2 of 8

Test method and its validation:

Method: Dilution-neutralisation

Neutraliser: 100.0 g/l Polysorbate 80 + 30.0 g/l Lecithin +
30.0 g/l Tryptone Soya Broth + 5.0 g/l Sodium thiosulphate +
1.0 g/l L-histidine (Neutraliser B)

Neutraliser validation: Validated in accordance with EN 13727:2012+A2:2015 (5.5.2)

Experimental conditions:

Period of analysis: 27 February 2019 to 1 March 2019

Product test concentration(s): 50%

Diluent used for product test solution(s): Hard water

Contact time(s): 30 s ± 5 s

Test temperature(s): 20°C ± 1°C

Interfering substance: 3.0 g/l bovine albumin + 3.0 ml/l sheep erythrocytes
(dirty conditions)

Temperature of incubation: 36°C ± 1°C

Identification of the bacterial strain(s) used: *Pseudomonas aeruginosa* (DSM 939)
Escherichia coli K12 (NCIMB 10083)
Staphylococcus aureus (DSM 799)
Enterococcus hirae (DSM 3320)

Deviations: None

Remarks:

- 1) All test conditions are as requested by the client, irrespective of whether these are in accordance with EN 13727:2012+A2:2015 (5.4.2) or EN 13727:2012+A2:2015 (5.5.1.1).
- 2) Handwash products are always pre-diluted with hard water. This solution simulates the addition of tap water in practice (1:1).

Requirements:

The product shall demonstrate at least a 3 decimal log (lg) reduction against all test organisms.

Conclusion:

According to EN 13727:2012+A2:2015, this sample of Handwash possesses bactericidal activity when tested at a concentration of 50% with a contact time of 30 seconds at 20°C under dirty conditions against all of the referenced strains of *Pseudomonas aeruginosa*, *Escherichia coli* K12, *Staphylococcus aureus* and *Enterococcus hirae*.

Report prepared by:

Signed:



Name:

Tony Watson

Position:

General Manager

Date:

5 March 2019

Approved by:

Signed:



Name:

Gareth Bayliss

Position:

Laboratory Manager

Date:

5 March 2019

Results: EN 13727:2012+A2:2015

Test organism: *Pseudomonas aeruginosa* (DSM 939)
 Date of test: 27 February 2019
 Test temperature: 20°C ± 1°C Incubation temperature: 36°C ± 1°C
 Dilution-neutralisation method: Pour plate Number of plates: 1 / ml
 Neutraliser: B Test conditions: Dirty conditions

Validation and controls:

Validation suspension (N_{V_0})			Experimental conditions control (A)			Neutraliser or filtration control (B)			Method validation (C) Product conc.: 50%		
Vc1	74	\bar{x} =	Vc1	71	\bar{x} =	Vc1	75	\bar{x} =	Vc1	74	\bar{x} =
Vc2	79	76.5	Vc2	73	72	Vc2	79	77	Vc2	74	74
30 ≤ \bar{x} of N_{V_0} ≤ 160 ? <input checked="" type="checkbox"/> yes <input type="checkbox"/> no			\bar{x} of A ≥ 0.5 x \bar{x} of N_{V_0} ? <input checked="" type="checkbox"/> yes <input type="checkbox"/> no			\bar{x} of B ≥ 0.5 x \bar{x} of N_{V_0} (or $N_{V_B} / 1000$) ? <input checked="" type="checkbox"/> yes <input type="checkbox"/> no			\bar{x} of C ≥ 0.5 x \bar{x} of N_{V_0} ? <input checked="" type="checkbox"/> yes <input type="checkbox"/> no		
Validation suspension (N_{V_B})			Vc1	76	\bar{x} =	30 ≤ \bar{x} of $N_{V_B} / 1000$ ≤ 160 ?			<input checked="" type="checkbox"/> yes <input type="checkbox"/> no		
			Vc2	79	77.5						

Test suspension (N and N_0):

N	Vc1	Vc2	\bar{x} wm = 4.40 x 10 ⁸ ;	lg N = 8.64
10 ⁻⁶	>330	>330	$N_0 = N / 10$;	lg N_0 = 7.64
10 ⁻⁷	42	46	7.17 ≤ lg N_0 ≤ 7.70 ?	<input checked="" type="checkbox"/> yes <input type="checkbox"/> no

Test:

Conc. of the product	Contact time	Dilution step	Vc1	Vc2	N_a (\bar{x} x 10 or \bar{x} wm x 10)	lg N_a	lg R (lg N_0 - lg N_a)
50%	30 s	10 ⁰	4	0	<1.40 x 10 ²	<2.15	>5.49
		10 ⁻¹	0	0			
		10 ⁻²	0	0			

Results: EN 13727:2012+A2:2015

Test organism:	<i>Escherichia coli</i> K12	(NCIMB 10083)
Date of test:	27 February 2019	
Test temperature:	20°C ± 1°C	Incubation temperature: 36°C ± 1°C
Dilution-neutralisation method:	Pour plate	Number of plates: 1 / ml
Neutraliser:	B	Test conditions: Dirty conditions

Validation and controls:

Validation suspension (N_{V_0})			Experimental conditions control (A)			Neutraliser or filtration control (B)			Method validation (C) Product conc.: 50%		
Vc1	70	\bar{x} =	Vc1	71	\bar{x} =	Vc1	64	\bar{x} =	Vc1	69	\bar{x} =
Vc2	68	69	Vc2	67	69	Vc2	66	65	Vc2	69	69
30 ≤ \bar{x} of N_{V_0} ≤ 160 ? <input checked="" type="checkbox"/> yes <input type="checkbox"/> no			\bar{x} of A ≥ 0.5 x \bar{x} of N_{V_0} ? <input checked="" type="checkbox"/> yes <input type="checkbox"/> no			\bar{x} of B ≥ 0.5 x \bar{x} of N_{V_0} (or $N_{V_B} / 1000$) ? <input checked="" type="checkbox"/> yes <input type="checkbox"/> no			\bar{x} of C ≥ 0.5 x \bar{x} of N_{V_0} ? <input checked="" type="checkbox"/> yes <input type="checkbox"/> no		
Validation suspension (N_{V_B})			Vc1	70	\bar{x} =	30 ≤ \bar{x} of $N_{V_B} / 1000$ ≤ 160 ?			<input checked="" type="checkbox"/> yes <input type="checkbox"/> no		
			Vc2	70	70						

Test suspension (N and N_0):

N	Vc1	Vc2	\bar{x} wm = 4.25 x 10 ⁸ ;	lg N = 8.63
10 ⁻⁶	>330	>330	$N_0 = N / 10$;	lg N_0 = 7.63
10 ⁻⁷	41	44	7.17 ≤ lg N_0 ≤ 7.70 ? <input checked="" type="checkbox"/> yes <input type="checkbox"/> no	

Test:

Conc. of the product	Contact time	Dilution step	Vc1	Vc2	N_a (\bar{x} x 10 or \bar{x} wm x 10)	lg N_a	lg R (lg N_0 - lg N_a)
50%	30 s	10 ⁰	0	0	<1.40 x 10 ²	<2.15	>5.48
		10 ⁻¹	0	0			
		10 ⁻²	0	0			

Results: EN 13727:2012+A2:2015

Test organism: *Staphylococcus aureus* (DSM 799)
 Date of test: 27 February 2019
 Test temperature: 20°C ± 1°C Incubation temperature: 36°C ± 1°C
 Dilution-neutralisation method: Pour plate Number of plates: 1 / ml
 Neutraliser: B Test conditions: Dirty conditions

Validation and controls:

Validation suspension (N_{V_0})			Experimental conditions control (A)			Neutraliser or filtration control (B)			Method validation (C) Product conc.: 50%		
Vc1	46	\bar{x} =	Vc1	53	\bar{x} =	Vc1	49	\bar{x} =	Vc1	49	\bar{x} =
Vc2	49	47.5	Vc2	50	51.5	Vc2	47	48	Vc2	53	51
30 ≤ \bar{x} of N_{V_0} ≤ 160 ? <input checked="" type="checkbox"/> yes <input type="checkbox"/> no			\bar{x} of A ≥ 0.5 x \bar{x} of N_{V_0} ? <input checked="" type="checkbox"/> yes <input type="checkbox"/> no			\bar{x} of B ≥ 0.5 x \bar{x} of N_{V_0} (or $N_{V_B} / 1000$) ? <input checked="" type="checkbox"/> yes <input type="checkbox"/> no			\bar{x} of C ≥ 0.5 x \bar{x} of N_{V_0} ? <input checked="" type="checkbox"/> yes <input type="checkbox"/> no		
Validation suspension (N_{V_B})			Vc1	48	\bar{x} =	30 ≤ \bar{x} of $N_{V_B} / 1000$ ≤ 160 ?			<input checked="" type="checkbox"/> yes <input type="checkbox"/> no		
			Vc2	47	47.5						

Test suspension (N and N_0):

N	Vc1	Vc2	\bar{x} wm = 2.25 x 10 ⁸ ;	lg N = 8.35
10 ⁻⁶	231	221	$N_0 = N / 10$;	lg N_0 = 7.35
10 ⁻⁷	23	21	7.17 ≤ lg N_0 ≤ 7.70 ?	<input checked="" type="checkbox"/> yes <input type="checkbox"/> no

Test:

Conc. of the product	Contact time	Dilution step	Vc1	Vc2	N_a (\bar{x} x 10 or \bar{x} wm x 10)	lg N_a	lg R (lg N_0 - lg N_a)
50%	30 s	10 ⁰	127	119	1.23 x 10 ³	3.09	4.26
		10 ⁻¹	12	13			
		10 ⁻²	1	2			

Results: EN 13727:2012+A2:2015

Test organism:	<i>Enterococcus hirae</i>	(DSM 3320)
Date of test:	27 February 2019	
Test temperature:	20°C ± 1°C	Incubation temperature: 36°C ± 1°C
Dilution-neutralisation method:	Pour plate	Number of plates: 1 / ml
Neutraliser:	B	Test conditions: Dirty conditions

Validation and controls:

Validation suspension (N_{V_0})			Experimental conditions control (A)			Neutraliser or filtration control (B)			Method validation (C) Product conc.: 50%		
Vc1	40	\bar{x} =	Vc1	41	\bar{x} =	Vc1	37	\bar{x} =	Vc1	49	\bar{x} =
Vc2	43	41.5	Vc2	40	40.5	Vc2	42	39.5	Vc2	37	43
30 ≤ \bar{x} of N_{V_0} ≤ 160 ? <input checked="" type="checkbox"/> yes <input type="checkbox"/> no			\bar{x} of A ≥ 0.5 x \bar{x} of N_{V_0} ? <input checked="" type="checkbox"/> yes <input type="checkbox"/> no			\bar{x} of B ≥ 0.5 x \bar{x} of N_{V_0} (or $N_{V_B} / 1000$) ? <input checked="" type="checkbox"/> yes <input type="checkbox"/> no			\bar{x} of C ≥ 0.5 x \bar{x} of N_{V_0} ? <input checked="" type="checkbox"/> yes <input type="checkbox"/> no		
Validation suspension (N_{V_B})			Vc1	39	\bar{x} =	30 ≤ \bar{x} of $N_{V_B} / 1000$ ≤ 160 ?			<input checked="" type="checkbox"/> yes <input type="checkbox"/> no		
			Vc2	37	38						

Test suspension (N and N_0):

N	Vc1	Vc2	\bar{x} wm = 3.01 x 10 ⁸ ;	lg N = 8.48
10 ⁻⁶	297	304	$N_0 = N / 10$;	lg N_0 = 7.48
10 ⁻⁷	29	33	7.17 ≤ lg N_0 ≤ 7.70 ?	<input checked="" type="checkbox"/> yes <input type="checkbox"/> no

Test:

Conc. of the product	Contact time	Dilution step	Vc1	Vc2	N_a (\bar{x} x 10 or \bar{x} wm x 10)	lg N_a	lg R (lg N_0 - lg N_a)
50%	30 s	10 ⁰	10	14	<1.40 x 10 ²	<2.15	>5.33
		10 ⁻¹	1	2			
		10 ⁻²	0	0			

Explanations:

- V_c count per ml (one plate or more)
- \bar{x} average of V_{c1} and V_{c2} (1 + 2 duplicate)
- \bar{x}_{wm} weighted mean of \bar{x}
- N number of cells per ml in the test suspension
- N_o number of cells in the test mixture at the beginning of the contact time ($N_o = N / 10$)
- N_a number of survivors per ml in the test mixture at the end of the contact time (before neutralisation or filtration)
- R reduction ($\lg R = \lg N_o - \lg N_a$)
- N_v number of cells per ml in the validation suspension
- N_{v_o} number of cells in the validation mixtures at the beginning of the contact time ($N_{v_o} = N_v / 10$)
- A number of survivors per ml in the experimental conditions control mixture at the end of the contact time
- B number of survivors per ml in the neutraliser or filtration control mixture after 5 minutes
- C number of survivors per ml in the method validation mixture after 30 minutes