

**Report:** SSL.19B055.MB3

**Issued:** 5 April 2019

**Page:** 1 of 7

**Test Report:**

**EN 13727:2012+A2:2015**

Chemical disinfectants and antiseptics – Quantitative suspension test for the evaluation of bactericidal activity in the medical area – Test method and requirements (phase 2, step 1)

**Identification of the test laboratory:**

Abbott Analytical Ltd  
Unit 2, Hickmans Road, Birkenhead, CH41 1JH, United Kingdom

**Identification of the client:**

Safe Solutions (Safe4) Ltd  
Wharton Green, Bostock Road, Winsford, CW7 3BD, United Kingdom

**Identification of the sample:**

19B/055

Name of the product:	Safe4 Instrument Disinfectant
Batch number/reference and expiry date (if available):	3060
Date of delivery:	25 February 2019
Storage conditions:	Room temperature in darkness
Product diluent recommended by the manufacturer for use:	Not disclosed
Active substance(s) and their concentrations (s) (optional):	Not disclosed
Appearance of the product:	Clear green liquid

**Notes:**

- 1) The test results in this report relate only to the sample(s) tested.
- 2) This test report may not be reproduced except in full, adapted, altered or used to create a derivative work, without written approval from Abbott Analytical Ltd.

**Report:** SSL.19B055.MB3

**Issued:** 5 April 2019

**Page:** 2 of 7

**Test method and its validation:**

**Method:** Dilution-neutralisation

**Neutraliser:** 30.0 g/l Polysorbate 80 + 3.0 g/l Lecithin + 1.0 g/l L-histidine +  
1.0 g/l L-cysteine (Neutraliser A)  
100.0 g/l Polysorbate 80 + 30.0 g/l Lecithin +  
30.0 g/l Tryptone Soya Broth + 5.0 g/l Sodium thiosulphate +  
1.0 g/l L-histidine (Neutraliser B)

**Neutraliser validation:** Validated in accordance with EN 13727:2012+A2:2015 (5.5.2)

**Experimental conditions:**

**Period of analysis:** 12 March 2019 to 4 April 2019

**Product test concentration(s):** 1:20

**Diluent used for product test solution(s):** Hard water

**Contact time(s):** 10 min  $\pm$  10 s

**Test temperature(s):** 20°C  $\pm$  1°C

**Interfering substance:** 0.3 g/l bovine albumin (clean conditions)

**Temperature of incubation:** 36°C  $\pm$  1°C

**Identification of the bacterial strain(s) used:** *Pseudomonas aeruginosa* (NCIMB 10421)  
*Staphylococcus aureus* (NCTC 10788)  
*Enterococcus hirae* (NCIMB 8192)

**Deviations:** None

**Remarks:**

- 1) All test conditions are as requested by the client, irrespective of whether these are in accordance with EN 13727:2012+A2:2015 (5.4.2) or EN 13727:2012+A2:2015 (5.5.1.1).

**Requirements:**

The product shall demonstrate at least a 5 decimal log (lg) reduction against every test organism.

**Conclusion:**

According to EN 13727:2012+A2:2015, Safe4 Instrument Disinfectant possesses bactericidal activity when tested at a concentration of 1:20 with a contact time of 10 minutes at 20°C under clean conditions against all of the referenced strains of *Pseudomonas aeruginosa*, *Staphylococcus aureus* and *Enterococcus hirae*.

**Report prepared by:**

Signed:



Name:

Tony Watson

Position:

General Manager

Date:

5 April 2019

**Approved by:**

Signed:



Name:

Gareth Bayliss

Position:

Laboratory Manager

Date:

5 April 2019

**Results:** EN 13727:2012+A2:2015

Test organism: *Pseudomonas aeruginosa* (NCIMB 10421)  
 Date of test: 12 March 2019  
 Test temperature: 20°C ± 1°C Incubation temperature: 36°C ± 1°C  
 Dilution-neutralisation method: Pour plate Number of plates: 1 / ml  
 Neutraliser: A Test conditions: Clean conditions

**Validation and controls:**

Validation suspension ( $N_{V_0}$ )			Experimental conditions control (A)			Neutraliser or filtration control (B)			Method validation (C) Product conc.: 1:20		
Vc1	50	$\bar{x}$ =	Vc1	41	$\bar{x}$ =	Vc1	52	$\bar{x}$ =	Vc1	53	$\bar{x}$ =
Vc2	54	52	Vc2	49	45	Vc2	54	53	Vc2	50	51.5
30 ≤ $\bar{x}$ of $N_{V_0}$ ≤ 160 ? <input checked="" type="checkbox"/> yes <input type="checkbox"/> no			$\bar{x}$ of A ≥ 0.5 x $\bar{x}$ of $N_{V_0}$ ? <input checked="" type="checkbox"/> yes <input type="checkbox"/> no			$\bar{x}$ of B ≥ 0.5 x $\bar{x}$ of $N_{V_0}$ (or $N_{V_B} / 1000$ ) ? <input checked="" type="checkbox"/> yes <input type="checkbox"/> no			$\bar{x}$ of C ≥ 0.5 x $\bar{x}$ of $N_{V_0}$ ? <input checked="" type="checkbox"/> yes <input type="checkbox"/> no		
Validation suspension ( $N_{V_B}$ )			Vc1	52	$\bar{x}$ =	30 ≤ $\bar{x}$ of $N_{V_B} / 1000$ ≤ 160 ?			<input checked="" type="checkbox"/> yes <input type="checkbox"/> no		
			Vc2	51	51.5						

**Test suspension ( $N$  and  $N_0$ ):**

$N$	Vc1	Vc2	$\bar{x}$ wm = $4.10 \times 10^8$ ;	$\lg N = 8.61$
$10^{-6}$	>330	>330	$N_0 = N / 10$ ;	$\lg N_0 = 7.61$
$10^{-7}$	39	43	7.17 ≤ $\lg N_0$ ≤ 7.70 ? <input checked="" type="checkbox"/> yes <input type="checkbox"/> no	

**Test:**

Conc. of the product	Contact time	Dilution step	Vc1	Vc2	$N_a$ ( $\bar{x} \times 10$ or $\bar{x}$ wm x 10)	$\lg N_a$	$\lg R$ ( $\lg N_0 - \lg N_a$ )
1:20	10 min	$10^0$	0	0	<140	<2.15	>5.46
		$10^{-1}$	0	0			

**Results:** EN 13727:2012+A2:2015

Test organism: *Staphylococcus aureus* (NCTC 10788)  
 Date of test: 2 April 2019  
 Test temperature: 20°C ± 1°C Incubation temperature: 36°C ± 1°C  
 Dilution-neutralisation method: Pour plate Number of plates: 1 / ml  
 Neutraliser: B Test conditions: Clean conditions

**Validation and controls:**

Validation suspension ( $N_{V_0}$ )			Experimental conditions control (A)			Neutraliser or filtration control (B)			Method validation (C) Product conc.: 1:20		
Vc1	70	$\bar{x}$ = 71	Vc1	70	$\bar{x}$ = 69	Vc1	65	$\bar{x}$ = 68.5	Vc1	64	$\bar{x}$ = 68.5
Vc2	72		Vc2	68		Vc2	72		Vc2	73	
30 ≤ $\bar{x}$ of $N_{V_0}$ ≤ 160 ? <input checked="" type="checkbox"/> yes <input type="checkbox"/> no			$\bar{x}$ of A ≥ 0.5 x $\bar{x}$ of $N_{V_0}$ ? <input checked="" type="checkbox"/> yes <input type="checkbox"/> no			$\bar{x}$ of B ≥ 0.5 x $\bar{x}$ of $N_{V_0}$ (or $N_{V_B} / 1000$ ) ? <input checked="" type="checkbox"/> yes <input type="checkbox"/> no			$\bar{x}$ of C ≥ 0.5 x $\bar{x}$ of $N_{V_0}$ ? <input checked="" type="checkbox"/> yes <input type="checkbox"/> no		
Validation suspension ( $N_{V_B}$ )			Vc1	69	$\bar{x}$ = 70.5	30 ≤ $\bar{x}$ of $N_{V_B} / 1000$ ≤ 160 ?			<input checked="" type="checkbox"/> yes <input type="checkbox"/> no		
			Vc2	72							

**Test suspension ( $N$  and  $N_0$ ):**

$N$	Vc1	Vc2	$\bar{x}$ wm = 3.23 x 10 <sup>8</sup> ; $N_0 = N / 10$ ; 7.17 ≤ lg $N_0$ ≤ 7.70 ?
10 <sup>-6</sup>	324	319	lg $N$ = 8.51 ; lg $N_0$ = 7.51
10 <sup>-7</sup>	32	35	<input checked="" type="checkbox"/> yes <input type="checkbox"/> no

**Test:**

Conc. of the product	Contact time	Dilution step	Vc1	Vc2	$N_a$ ( $\bar{x}$ x 10 or $\bar{x}$ wm x 10)	lg $N_a$	lg $R$ (lg $N_0$ - lg $N_a$ )
1:20	10 min	10 <sup>0</sup>	0	0	<140	<2.15	>5.36
		10 <sup>-1</sup>	0	0			

**Results:** EN 13727:2012+A2:2015

Test organism:	<i>Enterococcus hirae</i>	(NCIMB 8192)
Date of test:	2 April 2019	
Test temperature:	20°C ± 1°C	Incubation temperature: 36°C ± 1°C
Dilution-neutralisation method:	Pour plate	Number of plates: 1 / ml
Neutraliser:	B	Test conditions: Clean conditions

**Validation and controls:**

Validation suspension ( $N_{V_0}$ )			Experimental conditions control (A)			Neutraliser or filtration control (B)			Method validation (C) Product conc.: 1:20		
Vc1	40	$\bar{x}$ =	Vc1	39	$\bar{x}$ =	Vc1	45	$\bar{x}$ =	Vc1	44	$\bar{x}$ =
Vc2	43	41.5	Vc2	39	39	Vc2	37	41	Vc2	40	42
30 ≤ $\bar{x}$ of $N_{V_0}$ ≤ 160 ? <input checked="" type="checkbox"/> yes <input type="checkbox"/> no			$\bar{x}$ of A ≥ 0.5 x $\bar{x}$ of $N_{V_0}$ ? <input checked="" type="checkbox"/> yes <input type="checkbox"/> no			$\bar{x}$ of B ≥ 0.5 x $\bar{x}$ of $N_{V_0}$ (or $N_{V_B} / 1000$ ) ? <input checked="" type="checkbox"/> yes <input type="checkbox"/> no			$\bar{x}$ of C ≥ 0.5 x $\bar{x}$ of $N_{V_0}$ ? <input checked="" type="checkbox"/> yes <input type="checkbox"/> no		
Validation suspension ( $N_{V_B}$ )			Vc1	39	$\bar{x}$ =	30 ≤ $\bar{x}$ of $N_{V_B} / 1000$ ≤ 160 ?			<input checked="" type="checkbox"/> yes <input type="checkbox"/> no		
			Vc2	40	39.5						

**Test suspension ( $N$  and  $N_0$ ):**

$N$	Vc1	Vc2	$\bar{x}$ wm = 1.99 x 10 <sup>8</sup> ;	lg $N$ = 8.30
10 <sup>-6</sup>	198	199	$N_0 = N / 10$ ;	lg $N_0$ = 7.30
10 <sup>-7</sup>	21	19	7.17 ≤ lg $N_0$ ≤ 7.70 ?	<input checked="" type="checkbox"/> yes <input type="checkbox"/> no

**Test:**

Conc. of the product	Contact time	Dilution step	Vc1	Vc2	$N_a$ ( $\bar{x}$ x 10 or $\bar{x}$ wm x 10)	lg $N_a$	lg $R$ (lg $N_0$ - lg $N_a$ )
1:20	10 min	10 <sup>0</sup>	0	0	<140	<2.15	>5.15
		10 <sup>-1</sup>	0	0			

**Explanations:**

- $V_c$  count per ml (one plate or more)
- $\bar{x}$  average of  $V_{c1}$  and  $V_{c2}$  (1 + 2 duplicate)
- $\bar{x}_{wm}$  weighted mean of  $\bar{x}$
- $N$  number of cells per ml in the test suspension
- $N_o$  number of cells in the test mixture at the beginning of the contact time ( $N_o = N / 10$ )
- $N_a$  number of survivors per ml in the test mixture at the end of the contact time (before neutralisation or filtration)
- $R$  reduction ( $\lg R = \lg N_o - \lg N_a$ )
- $N_v$  number of cells per ml in the validation suspension
- $N_{v_o}$  number of cells in the validation mixtures at the beginning of the contact time ( $N_{v_o} = N_v / 10$ )
- $A$  number of survivors per ml in the experimental conditions control mixture at the end of the contact time
- $B$  number of survivors per ml in the neutraliser or filtration control mixture after 5 minutes
- $C$  number of survivors per ml in the method validation mixture after 30 minutes